

**SECTION 1 :****E-TENDER NOTICE**

**BRIHANMUMBAI MUNICIPAL CORPORATION**  
HINDU HRIDAY SAMRAT BALASAHEB THACKERAY TRAUMA CARE HOSPITAL  
JOGESHWARI EAST-400060

**e- PROCUREMENT TENDER NOTICE**

**No. HO/HBTH/ 2944 Dt. 11.10.2024**

**e-Tender ID-2024\_MCGM\_1112031**

The Commissioner of Brihanmumbai Municipal Corporation invites the following online tender. The tender copy can be downloaded from BMC's portal (<http://www.mcgm.gov.in>) under "Tenders" section. However, the bid will be invited through Mahatender portal (<https://mahatenders.gov.in>) only.

Bidders who wish to participate in the Bidding process must register on the website <http://www.mahatenders.gov.in/nicgep/app>. Bidders, whose registration is valid, may please ignore this step. At the time of enrolment, the information required for enrolment should be filled. After enrolment the bidder will get his user name and password to his Mail Id. Bidders should have valid Class III Digital Signature Certificate (DSC) obtained from any licensed Certifying Authorities (CA). For registration, enrolment for digital signature certificates and user manual, Interested Bidders should follow the respective links provided in Mahatenders Portal (<https://mahatenders.gov.in>)

All interested vendors, are required to be registered with BMC .Vendors not registered with BMC before can apply online by clicking the link „Vendor Registration“ under the „eProcurement“ section of BMC Portal, Vendors already registered with BMC need to contact helpdesk to extend their vendor registration. The administrative, technical and commercial bids shall be submitted online up to the end date & time mentioned below.

Sr. No	Description	Tender Fee (Rs.)	EMD (Rs.)	Start Date and Time of online Bid Downloading	End Date and Time Of online Bid Submission
1	Purchase of 3 Steps Slit Lamp with Table (1 nos) for ophthalmology OPD at HBT Trauma Care Hospital Jogeshwari with CMC for five years after the completion of warranty of three years. <b>e-Tender</b>  <b>ID-2024_MCGM_1112031</b>	1,320/- +18% GST	16000/-	11.10.2024 13.00 hrs	18.10.2024 13.00 hrs

Note: Last date for on line payment of Earnest Money Deposit (EMD) is, on or before due date and time prescribed.

The tender document is available on BMC portal (<http://www.BMC.gov.in>) along with this tender notice. However, the bid will be invited through Mahatenders portal (<https://mahatenders.gov.in>)

**Earnest Money Deposit (EMD) shall be paid on line through payment gateway on or before due date and time prescribed. The vendors having standing deposit shall also have to pay full EMD amount online.**

The Authority (BMC) shall not be liable for any omission, mistake or error in respect of any of the above or on account of any matter or thing arising out of or concerning or relating to the tender or the Bidding Process, including any error or mistake therein or in any information or data given by the Authority.

The Municipal Commissioner reserves the right to reject all or any of the e-Tender(s) without assigning any reason at any stage.

Bidders shall note that any corrigendum issued regarding this tender notice/tender will be published on the BMC portal and Mahatenders portal only. No corrigendum will be published in the local newspapers.

**By Order of the**

**Municipal commissioner  
Brihanmumbai Municipal Corporation**

**Sd/-  
Medical Superintendent  
(HBT Trauma Care Hospital, Jogeshwari(E))**

Address for Communication and Venue for opening of bid :  
Office of the Medical Superintendent  
Hinduhriday samrat Balasaheb Thackeray  
Trauma Care Municipal Hospital,  
Jogeshwari (E), Mumbai 400 060.  
e-mail: [hbthjog13@gmail.com](mailto:hbthjog13@gmail.com)

**For detailed tender document please scroll down**

<b>Section: 2</b>		<b>HEADER DATA</b>	
<b>E-Tender File No.</b>	HO/ 2944 /HBTH Dt.23.07.2024		
<b>Name of Organization</b>	Brihanmumbai Municipal Corporation, HBT Hospital , Jogeshwari East, Mumbai –60		
<b>Mahatender ID No.</b>	e-Tender ID-2024_MCGM_1112031		
<b>Subject</b>	Purchase of 3 Steps Slit Lamp with Table (1 nos) for ophthalmology OPD at HBT Trauma Care Hospital Jogeshwari with CMC for five years after the completion of warranty of three years.		
<b>Contract Period</b>	<b>8 years (3 years warrantee + 5 years CMC)</b>		
<b>Estimated Cost</b>	Rs. 7,78,957.50/-		
<b>Tenders fee of E-Tender</b>	Rs.1,320/- + 18%GST		
<b>Earnest Money Deposit</b>	Rs. 16000/-		
<b>Start date</b>	<b>11.10.2024</b>	<b>from:13.00 hrs</b>	
<b>End date &amp; time of Bid Submission</b>	<b>18.10.2024</b>	<b>to:13.00 hrs</b>	
<b>End date &amp; time for receipt of EMD</b>	18.10.2024	to:13.00 hrs	
<b>Opening of Packet “A”</b>	<b>As mentioned in Mahatender Portal</b>		
<b>Opening of Packet “B”</b>			
<b>Opening of Packet “C”</b>			
<b>Address for Communication</b>	Office of the Medical Superintendent Hinduhriday samrat Balasaheb Thackeray Trauma Care Municipal Hospital, Jogeshwari (E), Mumbai400 060. e-mail: <a href="mailto:hbthjog13@gmail.com">hbthjog13@gmail.com</a>		
<b>Venue for opening of Bid</b>	Online at above address.		

**Section 3:PREAMBLE**

The Brihanmumbai Municipal Corporation invites Tenders from the manufacturer (Indian or Foreign)

**or**

100% Indian subsidiary of foreign manufacturer duly registered in India / Subsidiary of principle Foreign Manufacturer duly registered in India / sister concern of Foreign manufacturer duly registered in India /Associate of Foreign manufacturer duly registered in India /joint venture of Foreign manufacturer duly registered in India / affiliate of Foreign manufacturer duly registered in India

**or**

Distributor /Dealer / Importer /Traders/agent appointed directly by foreign manufacturer for the supply, installation, testing and commissioning of 3 Steps Slit Lamp with Table (1 nos) for opthal Opd of Hospital along with three year warranty.

along with Standard accessories and CMC for five years after the completion of warranty of three years to the Municipal Hospitals in Greater Mumbai as per the specification attached separately with this document and as per the terms and conditions as mentioned herein and as per the provisions of the M.M.C. Act, 1888 as amended till date.

## Technical Specification of 3 Steps Slit Lamp

Warranty period	3 years
Stereo Microscope Type	Galilean
Magnification Changer	Drum rotation
Working Distance	100mm
Objective Lens Focal Length	f100mm 0
Eye pieces (Large Optics)	12.5x
Total Magnifications	10x, 16x, 25x
Real Field of view	27,16,11mm
Inter Papillary Distance	55 - 75mm
Diopter Adjustment Range	-8 D to + 8 D
<b>Common Specification</b>	
Slit image width	0 to 14mm Continuously Variable
Slit image Height	0 to 14mm Continuously Variable
Illumination Field Diameter in mm	0.2,1,3,4,6,10 &14mm
Slit Rotation	0° to 1.80°
Slit Inclination	5°,10°,15°,20°
Filters	Heat absorbing, green (Red Free), Cobalt blue
Illumination	12V - 30W Halogen Lamp/LED 3.4 V, 700MAH,5W
Operation	3 dimensional Single handed operation
Instrument base movement	99mm (in/out), 30 mm (up/down), 118 mm (left / right)
Chin rest movement range	120mm
<b>Type:</b>	
Up/down movement range	150mm (10")
Load lifting capacity range	100kg
<b>Electrical Specification</b>	
Power Input	-110-230 V, 50-60Hz
Fuse	3A slow blow
Power consumption	45 W max (Halogen) / 31 VA Max [LED]
Dimensions of Stand	600(L) X 410 (B) 670 to 820 (H) 820mm
Dimensions of table top	520 (L) X 400 (B)mm
Table Top thickness	25mm
Nett Weight	20 kg
Gross weight	27 kg

Sd/-  
Lecturer  
Department of Ophthalmic  
HBT Trauma Hospital, Jogeshwari (E)

Sd/-  
HOD  
Department of Ophthalmic  
HBT Trauma Hospital, Jogeshwari (E)

## **General Specifications :**

1. It should CE Certified by European Notified Body (under MDD) along with declaration of conformity or US FDA approved for offered equipment and accessories.

In case of CE (Class-I) following documents are required.

- a) Declaration of conformity by manufacturer or EU representative of Manufacturer for the quoted model.
- b) Documentary evidence regarding firm registered with EEA (European economic Area) Competent authority is required.

**Or** European Representative registered with EEA (EUROPEAN ECONOMIC AREA) Competent authority appointed by firm is required.

**Or** other documents like certificated from notified body along with declaration of conformity.

2. Demonstration of the quoted model is compulsory
3. Power supply : 230 V +/- 15%, 50 Hz +/- 3%
4. Service Training to MEC Engineers and Operational training to user department.
5. Operating and detail service manual should be provided.
6. Tropicalization : Operating Temperature : 40°C  
Storage Temperature : 60°C  
Relative humidity: Up to 90% non condensing.