

E- QUOTATION NOTICE

MUNICIPAL CORPORATION OF GREATER MUMBAI
HINDU HRIDAY SAMRAT BALASAHEB THACKERAY TRAUMA CARE HOSPITAL
JOGESHWARI EAST-400060

E-PROCUREMENT E-QUOTATION NOTICE

No. HO/ 2942/HBTH Dated: 23.07.2024

The Commissioner of Brihanmumbai Municipal Corporation invites the following online Quotation. The quotation copy can be downloaded from BMC's portal (<http://www.BMC.gov.in>) under "E-Quotation's" section. However, the bid will be invited through Mahatender portal (<https://mahatenders.gov.in>) only.

Bidders who wish to participate in the Bidding process must register on the website <http://www.mahatenders.gov.in/nicgep/app>. Bidders, whose registration is valid, may please ignore this step. At the time of enrolment, the information required for enrolment should be filled. After enrolment the bidder will get his user name and password to his Mail Id. Bidders should have valid Class III Digital Signature Certificate (DSC) obtained from any licensed Certifying Authorities (CA). For registration, enrolment for digital signature certificates and user manual, Interested Bidders should follow the respective links provided in Mahatenders Portal (<https://mahatenders.gov.in>)

All interested vendors, are required to be registered with BMC .Vendors not registered with BMC before can apply online by clicking the link „Vendor Registration“ under the „eProcurement“ section of BMC Portal, Vendors already registered with BMC need to contact helpdesk to extend their vendor registration.

The administrative, technical and commercial bids shall be submitted online up to the end date & time mentioned below.

Sr. No	Description	E-Quotation scrutiny Price	EMD Amount Rs.	Start date & Time for online Downloading	End date & Time for online Sybmission
1	Purchase of NST Machine (1no) for OBGY dept of HBT Hospital. e-Tender ID-2024_MCGM_1055768	Rs.300/- (+18% GST)	6000/-	24.07.2024 16.00 hrs	31.07.2024 16.00 hrs

Note:- Last date for on line payment of Earnest Money Deposit (EMD) is, on or before due date and time prescribed.

- Corrigendum if any will be published on Mahatender Portal website only.

The e-Quotation document is available on BMC portal (<http://www.mcgm.gov.in>) along with this Quotation notice. However, the bid will be invited through Mahatender portal (<https://mahatenders.gov.in>)

Earnest Money Deposit (EMD) shall be paid on line through payment gateway on or before due date and time prescribed. The vendors having standing deposit shall also have to pay full EMD amount online.

The Authority (BMC) shall not be liable for any omission, mistake or error in respect of any of the above or on account of any matter or thing arising out of or concerning or relating to the e-Quotation or the Bidding Process, including any error or mistake therein or in any information or data given by the Authority.

The Municipal Commissioner reserves the right to reject all or any of the e-Quotation(s) without assigning any reason at any stage.

e-Quotations shall note that any corrigendum issued regarding this e-Quotation notice will be published on the BMC portal only.

**By Order of the
Municipal commissioner
Brihanmumbai Municipal Corporation**

**Sd/-
Medical Superintendent
(HBT Trauma Care Hospital, Jogeshwari(E))**

Address for Communication and Venue for opening of bid :

Office of the Medical Superintendent
HinduhridayasamratBalasahebThackeray
TraumaCare Municipal Hospital,
Jogeshwari(E), Mumbai400 060.
e-mail: hbthjog13@gmail.com

For detailed quotation document please scroll down

HEADER DATA	
Name of Organization	Brihanmumbai Municipal Corporation, HBT Hospital , Jogeshwari East, Mumbai –60
E-Quotation File No.	HO/ 2942 /HBTH Dt. 19.07.2024
Mahatender ID No.	e-Quotation ID-2024_MCGM_1055768
Subject	Purchase of NST Machine (1no) for OBGY dept of HBT Hospital.
Contract Period	3 years warrantee
Estimated Cost	Rs. 2,99,040.00/- (incl.of GST)
E Quotation scrutiny fee	Rs.300/- + 18%GST
Earnest Money Deposit	Rs. 6000/-
Start date	24.07.2024 from 16.00 hrs
End date & time of Bid Submission	31.07.2024 to 16.00 hrs
End date & time for receipt of EMD	31.07.2024 to 16.00 hrs
Opening of Packet “A”	As mentioned in https://mahatenders.gov.in
Opening of Packet “B”	
Opening of Packet “C”	
Address for Communication	Office of the Medical Superintendent HinduhridaysamratBalasahebThackeray TraumaCare Municipal Hospital, Jogeshwari(E), Mumbai400 060. e-mail: hbthjog13@gmail.com
Venue for opening of Bid	Online at above address.

Technical Specification of NST Machine

1. The system should be Microprocessor based Fetal Monitor providing continuous monitoring of fetal heart rate (FHR) along with maternally sensed fetal activity during ante partum testing for NST (Non-Stress Test) and for intensive monitoring of active labour, with twin fetal monitoring facility at the same time.
2. Automatic detection of dual fetal movement.
3. The unit should have 3 channels FHR 1, FHR 2, UC.
4. Heart Rate counting Range: 30 to 240BPM.
5. Frequency - DC-0.5 HZ.
6. Ultrasound Frequency -0.985 MHz
7. Intensity: Less than 10m W/Sq cm.
8. Patient Database: It should store at least 150 patient data.
9. Printer should have Facility to print on inbuilt thermal printer.
10. Display:- 1) Display Minimum 7". 2) Actual FHR1 & FHR2 in BPM.
11. It should have a fetal movement print function.
12. Computer interface through RS232 connection
13. Units supplied should hold a unique Serial/Identification number.
14. Provide with all accessories.
15. Demonstration is compulsory
16. The Company should be CE Certified / USFDA Certificate
17. Three year comprehensive warranty should be given.
18. Authorization letter should be submitted.
19. User manual and proper training should be given.
20. Power Supply: 240Vac, 50/60Hz

Sd/-
Assistant Professor
Dept of OBGY
HBT Trauma Care Hospital

Sd/-
Medical Superintendent
HBT Trauma Care Hospital

General Specifications :

1. It should CE Certified by European Notified Body (under MDD) along with declaration of conformity or US FDA approved for offered equipment and accessories.
In case of CE (Class-I) following documents are required.
 - a) Declaration of conformity by manufacturer or EU representative of Manufacturer for

the quoted model.

- b) Documentary evidence regarding firm registered with EEA (European economic Area)
Competent authority is required.

Or European Representative registered with EEA (EUROPEAN ECONOMIC AREA)
Competent authority appointed by firm is required.

Or other documents like certificated from notified body along with declaration of conformity.

2. Demonstration of the quoted model is compulsory
3. Third party inspection certificate should be applied from port of origin of shipping of equipment (from the parent companies country of origin)
4. Service Training to MEC Engineers and Operational training to user department.
5. Operating and detail service manual should be provided.
6. Tropicalization : Operating Temperature : 40°C
Storage Temperature : 60°C
Relative humidity: Up to 90% non condensing.